



# New City Scales Ltd

**Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service**

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
Milton Keynes, Bucks MK17 0PD  
Tel: 01296 714477  
Fax: 01296 714483  
E-mail: support@newcityscales.co.uk  
Web Site: www.newcityscales.co.uk

CERTIFICATE SERIAL NO: 9178

DATE: 30<sup>th</sup> October 2018

## QUALITY ASSURANCE

## CERTIFICATE OF TEST

CUSTOMER: Mick George Ltd

ADDRESS: Kennet Landfill Site  
Kennet Hall Farm Ind  
Dane Hill Road  
Newmarket  
CB8 7QX


TYPE:  
Ian Fellows Lucid

SERIAL NO  
207538

CAPACITY:  
50,000kg

We certify that the above instrument/equipment has been subjected to our full test procedures, and conforms to the original manufacturer's specification and/or appropriate British, National or International Standard; and is in accordance with your stated order requirements.

We further certify that the above instrument/equipment has been tested using equipment and standards which are traceable to National Standards. By virtue of the Weights and Measures Act 1985.

SIGNED:   
For and on Behalf of New City Scales.

POSITION: *Director*

DATE: *30/10/18*





# New City Scales Ltd

Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
 Milton Keynes, Bucks MK17 0PD  
 Tel: 01296 714477  
 Fax: 01296 714483  
 E-mail: support@newcityscales.co.uk  
 Web Site: www.newcityscales.co.uk

## CALIBRATION RECORD

**TYPE:** Lucid

**MANUFACTURER:** Ian Fellows

**SERIAL:** 207538

**CAPACITY / INCREMENTS:** 50,000kg x 20kg

**ALLOWABLE TOLERANCE** + - 20kg

### TEST READINGS

CALIBRATION WEIGHT	READING	ERROR
ZERO	0	_____
2,000kg	2,000kg	_____
4,000kg	4,000kg	_____
6,000kg	6,000kg	_____
8,000kg	8,000kg	_____
10,000kg	10,000kg	_____
12,000kg	12,000kg	_____
14,000kg	14,000kg	_____
16,000kg	16,000kg	_____
18,000kg	18,000kg	_____
20,000kg	20,000kg	_____
Weights removed substituted deadweight	19,480kg	_____
29,480kg	29,480kg	_____
39,480kg	39,480kg	_____
43,360kg	43,360kg	_____

### Individual loadcell test (corners)

No 1.	No 2.	No 3.	No 4.
16,000kg	16,000kg	16,020kg	16,000kg

**CALIBRATION WEIGHTS USED:** Supplied by New City Scales: Cert No TR10062

**TESTED BY NEW CITY SCALES**

**SIGNED**.....

**DATED:** 30<sup>th</sup> October 2018

