



New City Scales Ltd

Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,
Milton Keynes, Bucks MK17 0PD
Tel: 01296 714477
Fax: 01296 714483
E-mail: support@newcityscales.co.uk
Web Site: www.newcityscales.co.uk

CERTIFICATE SERIAL NO: 9178

DATE: 29th October 2018

QUALITY ASSURANCE

CERTIFICATE OF TEST

CUSTOMER: Mick George Ltd

ADDRESS: Chatteris Road
Coln Fen
Somersham
PE28 3DN

TYPE:
Ian Fellows Lucid

SERIAL NO
82075

CAPACITY:
50,000kg

We certify that the above instrument/equipment has been subjected to our full test procedures, and conforms to the original manufacturer's specification and/or appropriate British, National or International Standard; and is in accordance with your stated order requirements.

We further certify that the above instrument/equipment has been tested using equipment and standards which are traceable to National Standards. By virtue of the Weights and Measures Act 1985.

SIGNED: 
For and on Behalf of New City Scales.

POSITION: *Director*

DATE: *29/10/18*





New City Scales Ltd

Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,
Milton Keynes, Bucks MK17 0PD
Tel: 01296 714477
Fax: 01296 714483
E-mail: support@newcityscales.co.uk
Web Site: www.newcityscales.co.uk

CALIBRATION RECORD

TYPE : Lucid

MANUFACTURER: Ian Fellows

SERIAL: 82075

CAPACITY / INCREMENTS: 50,000kg x 20kg

ALLOWABLE TOLERANCE: + - 20kg

TEST READINGS

CALIBRATION WEIGHT	READING	ERROR
ZERO	0	_____
2,000kg	2,000kg	_____
4,000kg	4,000kg	_____
6,000kg	6,000kg	_____
8,000kg	8,000kg	_____
10,000kg	10,000kg	_____
12,000kg	12,000kg	_____
14,000kg	14,000kg	_____
16,000kg	16,000kg	_____
18,000kg	18,000kg	_____
20,000kg	20,000kg	_____
Weights removed substituted deadweight	19,920kg	_____
29,920kg	29,920kg	_____
39,920kg	39,920kg	_____
44,380kg	44,380kg	_____

Individual loadcell test (corners)

No 1.	No 2.	No 3.	No 4.
16,000kg	16,000kg	16,000kg	16,000kg

CALIBRATION WEIGHTS USED: Supplied by New City Scales: Cert TR10062

TESTED BY NEW CITY SCALES

SIGNED.....

DATED: 29th October 2108

