



# New City Scales Ltd

**Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service**

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
Milton Keynes, Bucks MK17 0PD  
Tel: 01296 714477  
Fax: 01296 714483  
E-mail: support@newcityscales.co.uk  
Web Site: www.newcityscales.co.uk

CERTIFICATE SERIAL NO: 9175

DATE: 31<sup>st</sup> October 2018

## QUALITY ASSURANCE

### CERTIFICATE OF TEST

CUSTOMER: Mick George Ltd

ADDRESS: Riverside Ind Est  
Off Nursery Road  
Marsh Lane  
Boston  
PE21 7PJ


TYPE: Ian Fellows Lucid

SERIAL NO 200229

CAPACITY: 50,000kg

We certify that the above instrument/equipment has been subjected to our full test procedures, and conforms to the original manufacturer's specification and/or appropriate British, National or International Standard; and is in accordance with your stated order requirements.

We further certify that the above instrument/equipment has been tested using equipment and standards which are traceable to National Standards. By virtue of the Weights and Measures Act 1985.

SIGNED:   
For and on Behalf of New City Scales.

POSITION: *Director*

DATE: *31/10/18*





# New City Scales Ltd

Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
Milton Keynes, Bucks MK17 0PD  
Tel: 01296 714477  
Fax: 01296 714483  
E-mail: support@newcityscales.co.uk  
Web Site: www.newcityscales.co.uk

## CALIBRATION RECORD

**TYPE:** Lucid

**MANUFACTURER:** Ian Fellows

**SERIAL :** 200229

**CAPACITY / INCREMENTS:** 50,000kg x 20kg

**ALLOWABLE TOLERANCE** + - 20kg

### TEST READINGS

CALIBRATION WEIGHT	READING	ERROR
ZERO	0	_____
2,000kg	2,000kg	_____
4,000kg	4,000kg	_____
8,000kg	8,000kg	_____
10,000kg	10,000kg	_____
14,000kg	14,000kg	_____
16,000kg	16,000kg	_____
20,000kg	20,000kg	_____
Substitute deadload	19,960kg	_____
29,960kg	29,960kg	_____
39,960kg	39,960kg	_____
44,420kg	44,420kg	_____

### Individual loadcell test (corners)

No 1.	No 2.	No 3.	No 4.
16,020kg	16,000kg	16,000kg	16,020kg

**CALIBRATION WEIGHTS USED:** Supplied by New City Scales: Cert No TR10379

**TESTED BY NEW CITY SCALES**

**SIGNED**.....

**DATED:** 31<sup>st</sup> October 2018

