



# New City Scales Ltd

**Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service**

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
Milton Keynes, Bucks MK17 0PD  
Tel: 01296 714477  
Fax: 01296 714483  
E-mail: support@newcityscales.co.uk  
Web Site: www.newcityscales.co.uk

CERTIFICATE SERIAL NO: 9196

DATE: 8<sup>th</sup> December 2018

## QUALITY ASSURANCE

### CERTIFICATE OF TEST

CUSTOMER: Mick George Ltd

ADDRESS: Rushton Landfill Site  
Oakley Road  
Rushton  
Nr Kettering

TYPE:	SERIAL NO	CAPACITY:
Ian Fellows Lucid (in weighbridge)	153553	50,000kg
Ian Fellows Lucid (out weighbridge)	145113	50,000kg

We certify that the above instrument/equipment has been subjected to our full test procedures and conforms to the original manufacturer's specification and/or appropriate British, National or International Standard; and is in accordance with your stated order requirements.

We further certify that the above instrument/equipment has been tested using equipment and standards which are traceable to National Standards. By virtue of the Weights and Measures Act 1985.

SIGNED:   
For and on Behalf of New City Scales.

POSITION: *Director*

DATE: *8/12/18*





# New City Scales Ltd

**Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service**

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
Milton Keynes, Bucks MK17 0PD  
Tel: 01296 714477  
Fax: 01296 714483  
E-mail: support@newcityscales.co.uk  
Web Site: www.newcityscales.co.uk

## CALIBRATION RECORD

**TYPE:** Lucid (Out Weighbridge)

**MANUFACTURER:** Ian Fellows

**SERIAL:** 145113

**CAPACITY / INCREMENTS:** 50,000kg x 20kg

**ALLOWABLE TOLERANCE** + - 20kg

### TEST READINGS

CALIBRATION WEIGHT	READING	ERROR
ZERO	0	_____
4,000kg	4,000kg	_____
8,000kg	8,000kg	_____
10,000kg	10,000kg	_____
12,000kg	12,000kg	_____
14,000kg	14,000kg	_____
16,000kg	16,000kg	_____
20,000kg	20,000kg	_____
Weights removed substituted deadweight	19,380kg	_____
29,380kg	29,380kg	_____
39,380kg	39,380kg	_____
43,260kg	43,260kg	_____

### Individual loadcell test (corners)

No 1.	No 2.	No 3.	No 4.
16,020kg	16,000kg	16,000kg	16,000kg

**CALIBRATION WEIGHTS USED:** Supplied by New City Scales: Cert No TR10062

**TESTED BY NEW CITY SCALES**

**SIGNED**.....

**DATED: 8<sup>th</sup> December 2018**





# New City Scales Ltd

Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
Milton Keynes, Bucks MK17 0PD  
Tel: 01296 714477  
Fax: 01296 714483  
E-mail: support@newcityscales.co.uk  
Web Site: www.newcityscales.co.uk

## CALIBRATION RECORD

**TYPE:** Lucid (In Weighbridge)

**MANUFACTURER:** Ian Fellows

**SERIAL :** 153553

**CAPACITY / INCREMENTS:** 50,000kg x 20kg

**ALLOWABLE TOLERANCE** + - 20kg

### TEST READINGS


CALIBRATION WEIGHT	READING	ERROR
ZERO	0	_____
4,000kg	4,000kg	_____
8,000kg	8,000kg	_____
10,000kg	10,000kg	_____
12,000kg	12,000kg	_____
14,000kg	14,000kg	_____
16,000kg	16,000kg	_____
20,000kg	20,000kg	_____
Weights removed substituted deadweight	19,380kg	_____
29,380kg	29,380kg	_____
39,380kg	39,380kg	_____
43,260kg	43,260kg	_____

### Individual loadcell test (corners)

No 1.	No 2.	No 3.	No 4.
16,000kg	16,000kg	16,000kg	16,000kg

**CALIBRATION WEIGHTS USED:** Supplied by New City Scales: Cert No TR10062

**TESTED BY NEW CITY SCALES**

**SIGNED:** 

**DATED:** 8<sup>th</sup> December 2018

