



# New City Scales Ltd

**Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service**

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
 Milton Keynes, Bucks MK17 0PD  
 Tel: 01296 714477  
 Fax: 01296 714483  
 E-mail: support@newcityscales.co.uk  
 Web Site: www.newcityscales.co.uk

CERTIFICATE SERIAL NO: 8826

DATE: 16<sup>th</sup> December 2017

## QUALITY ASSURANCE

### CERTIFICATE OF TEST

CUSTOMER: Mick George Ltd

ADDRESS: Rushton Landfill Site  
 Oakley Road  
 Rushton  
 Nr Kettering

TYPE:	SERIAL NO	CAPACITY:
Ian Fellows Lucid (in weighbridge)	153553	50,000kg
Ian Fellows Lucid (out weighbridge)	145113	50,000kg

We certify that the above instrument/equipment has been subjected to our full test procedures and conforms to the original manufacturer's specification and/or appropriate British, National or International Standard; and is in accordance with your stated order requirements.

We further certify that the above instrument/equipment has been tested using equipment and standards which are traceable to National Standards. By virtue of the Weights and Measures Act 1985.

SIGNED: .....  
 For and on Behalf of New City Scales.

POSITION: *Q.A. Manager*.....

DATE: *16-12-17*.....



# New City Scales Ltd

Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
Milton Keynes, Bucks MK17 0PD  
Tel: 01296 714477  
Fax: 01296 714483  
E-mail: support@newcityscales.co.uk  
Web Site: www.newcityscales.co.uk

## CALIBRATION RECORD

**TYPE:** Lucid (Out Weighbridge)

**MANUFACTURER:** Ian Fellows

**SERIAL:** 145113

**CAPACITY / INCREMENTS:** 50,000kg x 20kg

**ALLOWABLE TOLERANCE** + - 20kg

### TEST READINGS

CALIBRATION WEIGHT	READING	ERROR
ZERO	0	_____
4,000kg	4,000kg	_____
8,000kg	8,000kg	_____
10,000kg	10,000kg	_____
12,000kg	12,000kg	_____
14,000kg	14,000kg	_____
16,000kg	16,000kg	_____
20,000kg	20,000kg	_____
Weights removed substituted deadweight	17,960kg	_____
27,960kg	27,960kg	_____
37,960kg	37,960kg	_____
41,860kg	41,860kg	_____

### Individual loadcell test (corners)

No 1.	No 2.	No 3.	No 4.	No 5.	No 6.
10,000kg	10,000kg	10,000kg	10,000kg	10,000kg	10,000kg

**CALIBRATION WEIGHTS USED:** Supplied by New City Scales: Cert No TR10063

TESTED BY NEW CITY SCALES

SIGNED.....

DATED: 16<sup>th</sup> December 2017



# New City Scales Ltd

Weighbridge Manufacturer, Sales, Hire, Calibration, Repairs, Service

Milton Keynes: Unit 3A, The Hollows, Winslow Road, Little Horwood,  
Milton Keynes, Bucks MK17 0PD  
Tel: 01296 714477  
Fax: 01296 714483  
E-mail: support@newcityscales.co.uk  
Web Site: www.newcityscales.co.uk

## CALIBRATION RECORD

**TYPE:** Lucid (In Weighbridge)

**MANUFACTURER:** Ian Fellows

**SERIAL :** 153553

**CAPACITY / INCREMENTS:** 50,000kg x 20kg

**ALLOWABLE TOLERANCE** + - 20kg

### TEST READINGS

CALIBRATION WEIGHT	READING	ERROR
ZERO	0	_____
4,000kg	4,000kg	_____
8,000kg	8,000kg	_____
10,000kg	10,000kg	_____
12,000kg	12,000kg	_____
14,000kg	14,000kg	_____
16,000kg	16,000kg	_____
20,000kg	20,000kg	_____
Weights removed substituted deadweight	17,960kg	_____
27,960kg	27,960kg	_____
37,960kg	37,960kg	_____
41,860kg	41,860kg	_____

### Individual loadcell test (corners)

No 1.	No 2.	No 3.	No 4.
16,000kg	16,000kg	16,000kg	16,000kg

**CALIBRATION WEIGHTS USED:** Supplied by New City Scales: Cert No TR10063

**TESTED BY NEW CITY SCALES**

**SIGNED:** 

**DATED:** 16<sup>th</sup> December 2017